



Certificate of Analysis

British Pharmacopoeia Phosphate Buffer pH 6.4

Lot N:	1079391
Description:	disodium hydrogen phosphate dodecahydrate R / sodium dihydrogen phosphate R / sodium chloride R / sodium hydroxide or 1 M hydrochloric acid / water R.
Ref N:	BPB 545
Performance test:	passes test
Measured value:	pH = 6.40
Expiry date:	23.02.2026

The reagent is specially intended for use in testing British Pharmacopoeia. This solution was produced according British Pharmacopoeia specifications using Ph reagent quality components and water meeting the Ph monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

This certificate relates solely to the lot number given above.

Certification date: 23.01.2025

Signed by: , Chemical Production Manager