



# Certificate of Analysis

## ***British Pharmacopoeia Lead Standard Solution (100 ppm Pb)***

Description:	Lead Standard Solution (0.1 per cent Pb)
Lot N:	1087148
Ref N:	BPS 419
Measured Concentration:	1000 +/- 5 ppm Pb
Method:	by ICP-AES
Expiry date:	14.03.2026

The reagent is specially intended for use in testing British Pharmacopoeia. This solution was produced according British Pharmacopoeia specifications using Ph reagent quality components and water meeting the Ph monograph requirement for Purified Water.

### *Instruction for use*

Immediately before use, dilute lead standard solution (0.1 per cent Pb) R to 10 times its volume with water R

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

This certificate relates solely to the lot number given above.

**Certification date:** 14.02.2025

Signed by: , Chemical Production Manager