



Certificate of Analysis

Chinese Pharmacopoeia Standard biphthalate BS

Lot N: 1139683
Content: 10.21g of potassium biphthalate in 1000ml water
Ref N: CM 27631
Performance Test: passed
Measured pH: 4.008 +/- 0.006 at 25C
Expiry date: 24.08.2026

The standard is specially intended for use in testing ChP and monographs. This solution was produced according to the specifications using ChP reagent quality components and water meeting the ChP monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions

This certificate relates solely to the lot number given above.

Certification date: 24.07.2025

Signed by: , Chemical Production Manager