



Certificate of Analysis

Chinese Pharmacopoeia Standard phosphate BS

Lot N:	1139684
Content:	3.55g of anhydrous disodium hydrogen phosphate and 3.40g of potassium dihydrogenphosphate in water to 1000ml
Ref N:	CM 27632
Performance Test:	passed
Measured pH:	6.868 +/- 0.008 at 25C
Expiry date:	24.08.2026

The standard is specially intended for use in testing ChP and monographs. This solution was produced according to the specifications using ChP reagent quality components and water meeting the ChP monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions

This certificate relates solely to the lot number given above.

Certification date: 24.07.2025

Signed by: , Chemical Production Manager