

## CERTIFIED REFERENCE MATERIAL Organic substance

Ref No: CM 30456  
Barcode: 41559536

Lot No: 1054188

Certification Date: 28.10.2024  
Expiry date: 28.11.2026

Description: Trichloroethene

CAS No: 79-01-6

Empirical formula: C<sub>2</sub>HCl<sub>3</sub>

MW: 131.388

**Certified Purity / Uncertainty:** 0.9860 +/- 0.0010 g/g (98.60 +/- 0.10 %)

Purity=100% - Assay organic impurities

**Water content:** 0.1

**Storage Conditions:** Store in a refrigerator at temperatures between 2°C to 8°C

Method of certification:	Concept of Certification and traceability statement:	Intended use:	Instructions for the correct use of this reference material:	Stability and storage:	Level of homogeneity:
CRM's calibration procedure (WQP 5.15.1/22)	The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with EA 4/02. Metrological traceability is established through in-house validated method. The measurement results are traceable to SI.	<b>For Laboratory Use Only.</b> This CRM is intended for: Calibration of TLC, GC/FID, GC/TCD, GC/ECD, GC/MS, GC/MS/MS, LC/UV, LC/MS and LC/MS/MS. Validation of analytical methods. Preparation of "working reference samples". Detection limit and linearity studies. This statement is not intended to restrict the use for other purposes.	This CRM can be used directly or can be diluted in an appropriate solvent. Only a clean glassware should be used. <b>Hazardous situation:</b> The normal laboratory safety precautions should be observed when working with this CRM. Further details for the handling of this chemical are available as safety data sheet.	This CRM is with a guaranteed purity +/- 2% deviation prior to the expiration date. Stability is guaranteed, provided that the material is kept in its original packaging, tightly closed stored, as written in the section: Storage Conditions.	The material was tested for homogeneity by analyzing randomly selected samples according to an in-house procedure. The level of homogeneity proved satisfactory for a sample volume of min. 2 mg. The uncertainty incorporates the sample standard deviation combined with the uncertainty calculated from homogeneity and stability studies.

This certificate relates solely to the lot number given above.

All processes (including generating of this certificate) are completely controlled by the specialized Computer-Aided-Manufacturing (CAM) software.

This Certified Reference Material was produced under a quality management system that is:

- Registered to ISO 9001 Quality Management System (Lloyd's Register Quality Assurance Ltd Cert No 0039638)
- Accredited according to ISO/IEC 17025
- Accredited according to ISO 17034

This document is designed and the certified value and uncertainty are determined in accordance with ISO 33401, ISO 33405, and Eurachem / CITAC Guides

### Names of certifying officers:

Laboratory:  Margarita Dimitrova

Manager:  Krassimira Taralova

### Analytical Data:

#### GC-Conditions

	He, constant flow	Oven	Temperature	Hold
Carrier gas:		Initial	70°C	0
Column:	Agilent CP9105 J&W VF-624ms 60m, 0.32mm, 1.80µm	10°C/min	120°C	0
Flow rate:	1.6 ml/min	5°C/min	160°C	0
Injection volume:	1µl split	30°C/min	220°C	3
Injector:	220 °C			

#### MS Conditions

Ionization mode:	EI
MS Quad:	150°C
MS Source:	230°C
Mode:	Scan
Transfer line:	230°C

Area Percent Report

Data Path : C:\DATA\2024\05.2024\  
Data File : Trichloroethene-41559536-s4.D  
Acq On : 7 May 2024 12:33  
Operator : MD  
Sample : Trichloroethene-41559536-s4  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Integration Parameters: autoint1.e  
Integrator: ChemStation

Method : C:\METHODS\Quant\995606.M  
Title :

Signal : TIC: Trichloroethene-41559536-s4.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.668	257	269	287	M3	43288	1538065	0.52%	0.513%
2	6.762	538	565	603	M	9340886	296151669	100.00%	98.707%
3	9.269	1227	1245	1271	M4	72488	2341674	0.79%	0.780%

Sum of corrected areas: 300031409

995606.M Tue May 07 13:09:45 2024

