



Certificate of Analysis

European Pharmacopoeia TOC water for Total organic carbon (TOC) determination

Lot N:	1142526
Description:	Ph Eur 2.2.44
Ref N:	CM 33018
Total organic carbon:	< 0.1 mg/l
Conductivity:	< 0.1 mS/cm at 25°C
Nitrates:	< 0.2 ppm
Aluminium:	< 10 ppb
Heavy metals:	< 0.1 ppm
Microbiological monitoring:	passes test
Expiry date:	06.09.2026

The reagent is specially intended for use in TOC determination of water for pharmaceutical use. It is not suitable for use in the manufacture of dialysis solutions. It is not sterile and apyrogenic.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

This certificate relates solely to the lot number given above.

Certification date: 06.08.2025

Signed by: , Chemical Production Manager