



Certificate of Analysis

Japanese Pharmacopoeia Standard Solution for limit tests Standard Cyanide Stock Solution

Lot N: 1021267
Ref N: JPHSS 074
Measured Concentration: 1.004 mg/ml CN⁻ +/- 0.5%
Method: Titration with 0.1 mol/L silver nitrate VS
Expiry date: 08.08.2025

The standard is specially intended for use in testing Japanese Pharmacopoeia and monographs. This solution was produced according to Japanese Pharmacopoeia specifications using reagent quality components and water meeting the Japanese Pharmacopoeia monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored in a refrigerator between 2 °C to 8 °C.

This certificate relates solely to the lot number given above.

Certification date: 08.07.2024

Signed by: , Chemical Production Manager