



Certificate of Analysis

Japanese Pharmacopoeia Potassium permanganate TS

Lot N: 1008470
Description: potassium permanganate, 0.02 mol/L
Ref N: JPHTS 436
Content: Potassium permanganate in water
Performance test: passed test
Expiry date: 29.06.2025

The standard is specially intended for use in testing Japanese Pharmacopoeia and monographs. This solution was produced according of Japanese Pharmacopoeia specifications using reagent quality components and water meeting the Japanese Pharmacopoeia monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

This certificate relates solely to the lot number given above.

Certification date: 29.05.2024

Signed by: , Chemical Production Manager