



Certificate of Analysis

European Pharmacopoeia Phosphomolybdotungstic reagent

Lot N:	1162526
Description:	Ph Eur N 1065000
Ref N:	RP 153 100
Starting materials:	sodium tungstate R / sodium molybdate R / hydrochloric acid R / phosphoric acid R / 150 g of lithium sulfate R / water R / bromine R.
Performance test:	passes test
Expiry date:	08.11.2026

The reagent is specially intended for use in testing Ph Eur and monographs. This solution was produced according 4.1.1 of Ph Eur specifications using Ph Eur reagent quality components and water meeting the Ph Eur monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored at 2 °C to 8 °C.

The reagent should be yellow in colour. If it acquires a greenish tint, it is unsatisfactory for use but may be regenerated by boiling with a few drops of bromine R. Care must be taken to remove the excess of bromine by boiling.

This certificate relates solely to the lot number given above.

Certification date: 08.10.2025

Signed by: , Chemical Production Manager