



# Certificate of Analysis

## ***European Pharmacopoeia Stannous chloride solution***

**Lot N:** 1144365  
**Description:** Ph Eur N 1085001  
**Ref N:** RP 205 100  
**Starting materials:** Tin R / hydrochloric acid R  
**Performance test:** passes test  
**Expiry date:** 07.09.2026

The reagent is specially intended for use in testing Ph Eur and monographs. This solution was produced according 4.1.1 of Ph Eur specifications using Ph Eur reagent quality components and water meeting the Ph Eur monograph requirement for Purified Water.

### *Instruction for use*

Prior to use add the tin shots (accompanying the bottle) to the solution. After tin shots addition to the main solution, the later should be used in no more than 48 hours.

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under ambient temp and protected from air.

This certificate relates solely to the lot number given above.

**Certification date:** 07.08.2025

Signed by: , Chemical Production Manager