



Certificate of Analysis

U.S. Pharmacopeia Hydrochloric Acid, Diluted (10 percent)

Lot N: 1098092
Description: hydrochloric acid / water
Ref N: USPS 362
Performance test: passes test (10.0% HCl)
Expiry date: 18.04.2026

The reagent is specially intended for use in testing USP and monographs. This solution was produced according USP specifications using USP reagent quality components and water meeting the USP monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

This certificate relates solely to the lot number given above.

Certification date: 18.03.2025

Signed by: , Chemical Production Manager